

MODERNWATER

# RaPID Assay<sup>®</sup>

## Atrazine

RaPID Assay<sup>®</sup> is a rapid field or laboratory enzyme immunoassay method for the analysis of soil and water for remediation, assessment, and industrial testing.

### Test result type

- Quantitative, semi-quantitative or qualitative depending on protocol selected

### Samples per kit

- 100 test kit (tests up to 80+ samples)

### Assay range

- 0.04 to 5.0 ppb
- LDD = 0.1 ppb

### Sample methods

- Methods exist for soil and a wide variety of commodities

### Sampling time

- 'Sample-to-Data' in approximately 60 minutes

- Rapid field testing procedure for analysis of soil, food or water samples
- Quantitative data results with excellent analytical precision
- Results available in approximately 60 minutes
- Tests up to 50 samples at once
- Magnetic particle immunoassay
- Training available
- EPA SW-846 method #4670 for drinking water



MODERNWATER



US: +1 302 669 6900  
UK: +44 (0)1483 696 030  
[www.modernwater.com](http://www.modernwater.com)  
[info@modernwater.co.uk](mailto:info@modernwater.co.uk)

## Specificity

The proprietary antibody used in this assay is specific for atrazine. The cross-reactivity of the Atrazine RaPID Assay® for various triazine analogues can be expressed as least detectable dose (LDD) which is estimated at 90% B/B<sub>0</sub> or as the concentration estimated (inhibitory concentration) at 50% B/B<sub>0</sub> (IC<sub>50</sub>).

ATRAZINE RAPID ASSAY (PPB)		
Contaminant	LDD	IC <sub>50</sub>
Atrazine	0.046	0.72
Ametryn	0.053	0.39
Propazine	0.033	0.74
Prometryn	0.054	0.64
Prometon	0.056	2.22
Simazine	0.340	4.90
Terbutryn	0.090	5.50
Terbutylazine	0.310	15.5
Desethyl Atrazine	0.062	3.21
Cyanazine	1.000	>10,000
6-Hydroxy Atrazine	1.100	148
Desisopropyl Atrazine	0.800	217

## Basic Test Procedure

- Add 200 µL of calibrator or prepared sample, 250µL enzyme conjugate and 500 uL antibody coupled magnetic particles. Vortex
- Incubate for 15 minutes
- Using the RaPID magnetic separator, decant and wash
- Add 500 µL colour solution
- Incubate for 20 minutes
- Stop the reaction and read colour at 450 nm
- Quantitative results and QC parameters are calculated and printed automatically using the RPA-II® spectrophotometer

## Test kit components

- Antibody coated magnetic particles, enzyme conjugate, colour development, stop and wash reagents for analysis of 100 test tubes
- Standards for 0.1 1.0 and 5.0 ppb Atrazine
- Zero standard and 3.0 ppb control standard
- Disposable test tubes
- Test kit instructions

## Storage & precautions

- Shelf life is typically one year from date of manufacture, with specific kit expiration date information provided on product packaging
- Reagents must be stored at 39° to 46°F (4° to 8°C) when not in use
- Store at ambient temperature 64° to 81°F (18° to 27°C) is acceptable for day of use
- Kits must be brought to 64° to 81°F (18° to 27°C) before use
- Do not expose colour solution to direct sunlight

## Required test materials

- Atrazine RaPID Assay
- Atrazine sample diluent (100 mL)

## Part #

A00071  
A00051

## Required test equipment

- Accessory kit (purchase) 6050100
- Accessory kit (rental) 6997010
- Accessory kit components:
  - Magnetic separation rack A00004
  - Repeator pipette A00008
  - Adjustable volume pipette A00176
  - RPA-II analyser with printer 6000111
  - Vortex mixer A00014
  - Digital timer A00015



MODERNWATER